

Ref:NCB/CA19-163/192009

January 13, 2020

ENQUIRY

=> By Speed Post/Courier

Dear Sirs

Please let us have your Quotation for the following:

S.No.	Item Description	Qty	UOM
1	<p>Drybath Incubator.</p> <p>The system should have LCD Display to view the set and actual parameters.</p> <p>The system should display time, speed and temperature on LCD Display.</p> <p>The system should have temperature ranging from 0°C to 100°C.</p> <p>The system should have Dual Over-heating protector for safety.</p> <p>The system should have time ranging from 1 min to 99 Hours.</p> <p>The system should have temperature control accuracy of $\pm 0.5^\circ\text{C}$.</p> <p>The system should have speed ranging from min of 100 RPM to max of 1500 RPM.</p> <p>The system should have heating time taken not more than 15 minutes.</p> <p>The system should have cooling time taken not more than 15 minutes.</p> <p>The system should have mixing orbit of min 2mm.</p> <p>The system should be supplied with Block 96 x 0.2 ml tubes, Block 54 x 0.5 ml, Block 35 x 1.5 ml tubes, Block 35 x 2.0 ml tubes.</p> <p>The system should be supplied with all the accessories required to function.</p> <p>The product should be as per CE/IEC guideline and certificate from authorized body should be submitted. No self-declaration will be accepted.</p> <p>Compliance to each of the above points should be separately indicated and evidence presence for each of them (Product brochures should be highlighted whenever required).</p> <p>The service team should be stationed in Bangalore. Office address and contact details of Engineer should be submitted with this offer.</p> <p>The past performance/ Service support in NCBS. inStem. CCamp should be satisfactory and the technical evaluation will be done accordingly.</p> <p>The system should have warranty period of 3 years.</p>	1.00	NO
2	<p>Drybath Incubator Shaker.</p> <p>The system should have LCD Display to view the set and actual parameters.</p> <p>The system should display time and temperature on LCD Display.</p> <p>The system should have easy method to replace the metal blocks.</p> <p>The system should have temperature ranging from -10°C to 100°C</p> <p>The system should have time ranging from 1 min to 99 Hours.</p> <p>The system should have temperature control accuracy of $\pm 0.5^\circ\text{C}$.</p> <p>The system should have heating time taken not more than 15 minutes.</p> <p>The system should have cooling time taken not more than 15 minutes.</p> <p>The system should be supplied with Block 96 x 0.2 ml tubes, Block 54 x 0.5 ml tubes, Block 35 x 1.5 ml tubes, Block 35 x 2 ml tubes.</p> <p>The system should be supplied with all the accessories required to function.</p> <p>The product should be as per CE/IEC guideline and certificate from authorized body should be submitted.</p> <p>Compliance to each of the above points should be separately indicated and evidence presence for each of them (Product brochures should be highlighted whenever required).</p> <p>The service team should be stationed in Bangalore. Office address and contact details of Engineer should be submitted with this offer.</p> <p>The past performance/ Service support in NCBS. inStem. CCamp should be satisfactory and the technical evaluation will be done accordingly.</p>	1.00	NO

S.No. Item Description	Qty	UOM
------------------------	-----	-----

The system should have warranty period of 3 years.

1. The quotation shall be submitted in a sealed envelope duly superscribed with the enquiry number, and the due date for Ref No: 192009. The bids should be addressed and to be mailed/couriered (sent by post/courier) to 'THE HEAD-PURCHASE'.

The bids are liable to be rejected if the sealed envelope is not addressed to "THE HEAD-PURCHASE" with Tender Ref No. and Item Description and due date. The bids delivered in person shall be dropped in Purchase Section. If the bids are sent through courier or mail, it should reach by submission date and time and NCBS will not be responsible for the delay.

2. DUE DATE FOR SUBMISSION OF QUOTATION AGAINST THIS ENQUIRY IS 31/01/2020 BY 5.30PM.

3. QUOTATIONS RECEIVED AFTER THE DUE DATE SHALL BE REJECTED.

4. The validity of your quotation should be for 60 days from the due date.

5. All duties, taxes, surcharge and cess as currently applicable must be stated in your quotation, separately. Otherwise your quote is liable to be rejected.

6. Your quotation should indicate delivery period & warranty period.

7. Delivery to be made to our stores. Please indicate charges, if any extra. Transit Insurance should be done upto NCBS Stores.

8. If you are unable to supply the quality, specifications or brand as mentioned in our enquiry, please state so and then offer alternative to quality/specifications.

9. Payment: within one month after delivery & acceptance/satisfactory installation.

10. Please ensure that the enquiry number and the due date is superscribed on the envelope failing which your quotation is liable to be rejected.

11. Since we are a public funded research institution, we are exempted from paying Customs duty (Except ad valorem duty of 5% + 2% Cess and 1% Cus Sec & High Edu. CESS vide Notification No. 51/96 with latest amendments) and excise duty vide Notification No. 10/97 CENTRAL EXCISE dated 01-03-1997 for all scientific equipments, technical instruments, equipments (including computers), their accessories, spares, consumables and software. Hence, please offer your prices

12. If the item is covered under DGS&D rate contract, please quote the rate as per the DGS&D rate contract with xerox copy of the DGS&D order.

13. Any dispute or differences that may arise between the parties shall be referred to the sole arbitration of the Centre Director or his nominees. The decision of the arbitrator shall be final and binding on the parties. The venue for arbitration shall be Bangalore. The provisions of the Arbitration and Conciliation Act, 1996 as amended from time to time shall apply. The courts in Bangalore shall have exclusive jurisdiction to deal with any or all disputes between the parties.


14. TIFR is a public funded research institute and is entitled to concessional rate of GST @ 5% for certain items supplied for reresearch purpose vide notification no. 47/2017 and 45/2017 dated 14th Nov, 2017. The offer should be submitted after fully considering the above notification.

15. Liquidity Damages: If the equipment/ items as per specifications in our P.O. is not supplied (shipped) within the specified delivery schedule, then liquidated damages (not in terms of penalty) will be imposed automatically and shall be deducted from the bill at the rate of 0.5% per week subject to a maximum of 10% of the order value.

16. Income Tax at the applicable rates as per the Indian Income Tax Act 1961 will be deducted at source for the services availed / ordered. In case of service provider, the rate of tax deduction shall be at 2% as per Section 194C, and in case of fee for professional / technical services under section 194J, the tax reduction shall be at the rate of 10%. In case service provider does not provide PAN number, the deduction shall be at 20% under section 206 AA.

-Tax Deduction Certificates will be provided on request for non PAN holders & Foreign Vendors and PAN holders could avail them through NSDL site dealing with 26AS.

Yours faithfully
For and on behalf of
National Centre for Biological Sciences

for 
Head Purchase