



National Center for Biological Sciences

Tata Institute of Fundamental Research

GKVK, Bellary Road, Bangalore 560 065. India

Tel. No : 080 2366, , 6346

Email: vinuthaks@ncbs.res.in, www.ncbs.res.in

Ref:NCB/CM17-1356/163598

February 12, 2018

ENQUIRY

=> By Speed Post/Courier

Dear Sirs

Please let us have your Quotation for the following:

S.No.	Item Description	Qty	UOM
1	Nitrile Rubber Tubing(pipe), closed cell of size ID - 1,1/8", Thickness - 1/2" (28.5mm x 12.7mm). As per the attached technical specifications. Preffered make: Aramcell,Superton,Supreme	200.00	MTR
2	Nitrile Rubber Tubing(pipe), closed cell of size ID - 1,3/8", Thickness - 3/4" (34.9mm x 19mm). As per the attached technical specifications. Preffered make: Aramcell,Superton,Supreme	300.00	MTR

NOTE :
1.Please provide AUTHORIZATION LETTER/PROPRIETARY CERTIFICATE along with quote 2. As per notification 45/2017 only 5% GST applicable to NCBS

1. The bids shall be enclosed in an envelope and due date sealed duly marked Tender for _____ Ref No: 163598. The bids should be addressed and to be mailed to 'THE HEAD-PURCHASE'.

The bids are liable to be rejected if the sealed envelope is not addressed to "THE HEAD-PURCHASE" with Tender Ref No. and Item Description and due date. The bids delivered in person shall be dropped in Purchase Section. If the bids are sent through courier or mail, it should reach by submission date and time and NCBS will not be responsible for the delay.

2. DUE DATE FOR SUBMISSION OF QUOTATION AGAINST THIS ENQUIRY IS 05/03/2018

3. QUOTATIONS RECEIVED AFTER THE DUE DATE SHALL BE REJECTED.

4. The validity of your quotation should be for 60 days from the due date.

5. All duties, taxes, surcharge and cess as currently applicable must be stated in your quotation, separately. Otherwise your quote is liable to be rejected.

6. Your quotation should indicate delivery period & warranty period.

7. Delivery to be made to our stores. Please indicate charges, if any extra. Transit Insurance should be done upto NCBS Stores.

8. If you are unable to supply the quality, specifications or brand as mentioned in our enquiry, please state so and then offer alternative to quality/specifications.

9. Payment: within one month after delivery & acceptance/satisfactory installation.

10. Please ensure that the enquiry number and the due date is superscribed on the envelope failing which your quotation is liable to be rejected.

11. Since we are a public funded research institution, we are exempted from paying Customs duty (Except ad valorem duty of 5% + 2% Cess and 1% Cus Sec & High Edu. CESS vide Notification No.51/96 with latest amendments) and excise duty vide Notification No. 10/97 CENTRAL EXCISE dated 01-03-1997 for all scientific equipments, technical instruments, equipments (including computers), their accessories, spares, consumables and software. Hence, please offer your prices

12. If the item is covered under DGS&D rate contract, please quote the rate as per the DGS&D rate contract with xerox copy of the DGS&D order.

13. Any dispute or differences that may arise between the parties shall be referred to the sole arbitration of the Centre Director or his nominees. The decision of the arbitrator shall be final and binding on the parties. The venue for arbitration shall be Bangalore. The provisions of the Arbitration and Conciliation Act, 1996 as amended from time to time shall apply. The courts in Bangalore shall have exclusive jurisdiction to deal with any or all disputes between the parties

Yours faithfully

For and on behalf of

National Centre for Biological Sciences

VINUTHA K S



**National Center for Biological Sciences
Specifications for Closed Cell Nitrile Rubber tubing**

AVERAGE PHYSICAL PROPERTIES OF INSULATION				
Cell structure	Closed cell			
	Minimum of 60Kg/M3 +/- 20Kg			
Thermal Conductivity(W/mK)	10°C	20° C	30°C	40°C
	0.0351	0.0362	0.0373	0.0385
Service Temperature Limit	(-20 ° C to 105° C)			
Ozone Resistance	Should be Excellent			
Flammability	Class O			
Resistance to chemicals	Organic Solvents	Should be Excellent		
	Dilute Inorganic acids	Should be very good		
	Mineral Oil	Should be very good		
Health aspects	should be Dust & Fibre free			
Water Absorption	Non hygroscopic coating			
Technical Broucher of the tubing offered to be submitted along with the quote				