



National Center for Biological Sciences
Tata Institute of Fundamental Research

GKVK, Bellary Road, Bangalore 560 065. India

Tel. No : 080 2366, 0

Email: shabeentaj@ncbs.res.in, www.ncbs.res.in

ENQUIRY

Ref: NCB/CA17-109/162499

October 16, 2017

Dear Sirs

Please let us have your Quotation for the following:

S.No.	Item Code	Description	Qty	UOM
1	LCQFR000022	-20° Freezer (As per attached specification)	1.00	NO

1. The bids shall be enclosed in an envelope and due date sealed duly marked Tender for _____ Ref No: 162499. The bids should be addressed and to be mailed to 'THE HEAD-PURCHASE'.

The bids are liable to be rejected if the sealed envelope is not addressed to "THE HEAD-PURCHASE" with Tender Ref No. and Item Description and due date. The bids delivered in person shall be dropped in Purchase Section. If the bids are sent through courier or mail, it should reach by submission date and time and NCBS will not be responsible for the delay.

2. DUE DATE FOR SUBMISSION OF QUOTATION AGAINST THIS ENQUIRY IS 06/11/2017.

3. QUOTATIONS RECEIVED AFTER THE DUE DATE SHALL BE REJECTED.

4. The validity of your quotation should be for 60 days from the due date.

5. All duties, taxes, surcharge and cess as currently applicable must be stated in your quotation, separately. Otherwise your quote is liable to be rejected.

6. Your quotation should indicate delivery period & warranty period.

7. Delivery to be made to our stores. Please indicate charges, if any extra. Transit Insurance should be done upto NCBS Stores.

8. If you are unable to supply the quality, specifications or brand as mentioned in our enquiry, please state so and then offer alternative to quality/specifications.

9. Payment: within one month after delivery & acceptance/satisfactory installation.

10. Please ensure that the enquiry number and the due date is superscribed on the envelope failing which your quotation is liable to be rejected.

11. Since we are a public funded research institution, we are exempted from paying Customs duty (Except advalorem duty of 5% + 2% Cess and 1% Cus Sec & High Edu. CESS vide Notification No. 51/96 with latest amendments) and excise duty vide Notification No. 10/97 CENTRAL EXCISE dated 01-03-1997 for all scientific equipments, technical instruments, equipments (including computers), their accessories, spares, consumables and software. Hence, please offer your prices

12. If the item is covered under DGS&D rate contract, please quote the rate as per the DGS&D rate contract with xerox copy of the DGS&D order.

13. Any dispute or differences that may arise between the parties shall be referred to the sole arbitration of the Centre Director or his nominees. The decision of the arbitrator shall be final and binding on the parties. The venue for arbitration shall be Bangalore. The provisions of the Arbitration and Conciliation Act, 1996 as amended from time to time shall apply. The courts in Bangalore shall have exclusive jurisdiction to deal with any or all disputes between the parties

Yours faithfully

For and on behalf of

National Centre for Biological Sciences



SHABEEN TAJ

Copy to: Prof. / Dr. / Mr. / Mrs. / Ms. / Shri / Smt. PARTH K RAVAL (NCB-12P4166)
Your Indent Req. No. NST4640033 Dated: 12/10/2017 refers.

-20° Freezer – 162 499

1. System should be vertical standing and should have working volume of 450-500 Ltrs
2. The system should be a research grade and suitable for continuous operation.
3. The system should have manual defrost.
4. The freezer should have double door.
5. External dimensions should not exceed 850x800x1800
6. External cabinet should be painted steel, internal should be styrol plastic
7. It should have Rigid polyurethane foamed-in place insulation.
8. The operating temperature for the system should be -20°C.
9. The system should have programmable temperature range of -10 to -30°C.
10. The temperature accuracy should be $\pm 1^{\circ}\text{C}$.
11. The system should have drawers for storage of sample and easy access.
12. The freezer should have audio and visual alarms for low/ high temperature, door open and power failure. Preferably there should be an option to mute the audible alarms.
13. The system should have programmable high/low temperature alarm.
14. The system should have microprocessor based temperature controller.
15. The system should have on board diagnosis.
16. It should be supplied with stabilizer and all other accessories required for proper functioning of freezer.
17. The system should have Caster moving wheels with stopper for easy movement
18. Must be offered with minimum 5year warranty
19. Power supply should be 230V, 50HZ
20. The system should have efficiency and low power consumption.
21. The system should be high power efficient and have high grade thermal insulation
22. Power consumption for 24 hrs at normal room condition must be mentioned
23. The system should have LED display for temperature.
24. The system should be certified for research grade as per norms of IS | CE or equivalent.
25. The Company must quote the freezer with all the suitable documents to justify the specification of product.
26. System should be compatible and have small foot print.
27. The company should have the service support in Bangalore