



National Center for Biological Sciences

Tata Institute of Fundamental Research

GKVK, Bellary Road, Bangalore 560 065. India

Tel. No : 080 2366, 6347

Email: shabeentaj@ncbs.res.in, www.ncbs.res.in

ENQUIRY

Ref: NCB/CA17-115/162553

October 24, 2017

Dear Sirs

Please let us have your Quotation for the following:

S.No.	Item Code	Description	Qty	UOM
1	LCQLI000026	CO2 Incubator(As per the specification attached)	1.00	NO

1. The bids shall be enclosed in an envelope and due date sealed duly marked Tender for _____ Ref No: 162553. The bids should be addressed and to be mailed to 'THE HEAD-PURCHASE'.

The bids are liable to be rejected if the sealed envelope is not addressed to "THE HEAD-PURCHASE" with Tender Ref No. and Item Description and due date. The bids delivered in person shall be dropped in Purchase Section. If the bids are sent through courier or mail, it should reach by submission date and time and NCBS will not be responsible for the delay.

2. DUE DATE FOR SUBMISSION OF QUOTATION AGAINST THIS ENQUIRY IS 14/11/2017.

3. QUOTATIONS RECEIVED AFTER THE DUE DATE SHALL BE REJECTED.

4. The validity of your quotation should be for 60 days from the due date.

5. All duties, taxes, surcharge and cess as currently applicable must be stated in your quotation, separately. Otherwise your quote is liable to be rejected.

6. Your quotation should indicate delivery period & warranty period.

7. Delivery to be made to our stores. Please indicate charges, if any extra. Transit Insurance should be done upto NCBS Stores.

8. If you are unable to supply the quality, specifications or brand as mentioned in our enquiry, please state so and then offer alternative to quality/specifications.

9. Payment: within one month after delivery & acceptance/satisfactory installation.

10. Please ensure that the enquiry number and the due date is superscribed on the envelope failing which your quotation is liable to be rejected.

11. Since we are a public funded research institution, we are exempted from paying Customs duty (Except ad valorem duty of 5% + 2% Cess and 1% Cus Sec & High Edu. CESS vide Notification No. 51/96 with latest amendments) and excise duty vide Notification No. 10/97 CENTRAL EXCISE dated 01-03-1997 for all scientific equipments, technical instruments, equipments (including computers), their accessories, spares, consumables and software. Hence, please offer your prices

12. If the item is covered under DGS&D rate contract, please quote the rate as per the DGS&D rate contract with xerox copy of the DGS&D order.

13. Any dispute or differences that may arise between the parties shall be referred to the sole arbitration of the Centre Director or his nominees. The decision of the arbitrator shall be final and binding on the parties. The venue for arbitration shall be Bangalore. The provisions of the Arbitration and Conciliation Act, 1996 as amended from time to time shall apply. The courts in Bangalore shall have exclusive jurisdiction to deal with any or all disputes between the parties

Yours faithfully

For and on behalf of

National Centre for Biological Sciences

Shabeen Taj

SHABEEN TAJ

Copy to: Prof. / Dr. / Mr. / Mrs. / Ms. / Shri / Smt. JAYAPRAKASH R. (NCB-12P4159)
Your Indent Req. No. JRA1200094 Dated: 17/10/2017 refers.

Remarks: CO2 Incubator

Specifications for Air jacketed CO2 incubator

Indent No: 162553/17-18

1. The system should be Air jacketed CO2 incubator
2. The system should be microprocessor controlled CO2 incubator It should have touch screen LCD controller with easy icons and user-friendly control system.
3. The system should have the Capacity of 170 Lt ($\pm 5\%$)
4. The system should be supplied with 3 nos of adjustable SS shelves
5. The system should have two doors and inner door should be glass door to seal the chamber with Silicon gasket.
6. The fluctuation/variation at 37°C should not be more than 0.1°C
7. The system should have the Infrared CO2 sensor
8. Sensor must be able to withstand high temperatures during direct heat sterilization
9. The system should have the CO2 range 1-20% with accuracy 0.1%
10. The system should have Built in on-demand high temperature decontamination cycle (180°C and above)
11. The system should have the option to mute the audible alarms
12. The system should have the tracking alarm $\pm 1^\circ \text{C}$
13. The system should have the seamless SS inner chamber.
14. The system should have the Operating temperature range from ambient to 50°C. The temperature accuracy and uniformity should be less than 0.1 °C
15. The system should have onboard Diagnostic interface to show system parameters and functions.
16. The system should be equipped with buzzer and visual alarm for power failure, CO2, and temperature deviation against the setup temperature
17. The incubator should be quoted with stackable kit with the main offer.
18. The system should operate at 230V, 50 Hz AC power
19. The system should be supplied with all the accessories required to function.
20. Compliance to each of the above points should be separately indicated and evidence presence for each of them (Product brochures should be highlighted wherever required).
21. The system should be supplied with 3-year warranty as a part of the offer.
22. The System should be CE/ISO Approved