

**ENQUIRY-Corrigendum 1**

Ref: NCB/CA19-168/192017 (V)

January 20, 2020  
=>By Speed Post/Courier

Dear Sirs,  
Please let us have your Quotation for the following :

Sl.No	Item Description	Qty
1	<p><b>Analytical Weighing Balance</b></p> <p><b>Specifications:</b></p> <p>The system should have Maximum Weighing capacity of 150g</p> <p>The system should have LCD Display with the attractive backlight.</p> <p>The system should show minimum display of 0.0001 g</p> <p>The system should have repeatability of 0.0001 g</p> <p>The system should have "Tare" function available.</p> <p>The system should have "Hold" function available.</p> <p>The system should have round weighing platform, 3-door draft shield glass chamber providing 8 inches height from pan to top of glass door.</p> <p>The system should have MEMORY function to keep the last weight in memory.</p> <p>The system should have inbuilt rechargeable battery backup for minimum of 8 hrs.</p> <p>The system have Internal Calibration option.</p> <p>The system should have response time of min 3 seconds.</p> <p>The system should be compatible in operating at 4°C to 40°C</p> <p>The system should be able to operate at 230V, 50 Hz.</p> <p>The system have option for power saving as a inbuilt feature.</p> <p>The system should have Low Battery indication should be available while operating on battery.</p> <p>The reading should get locked automatically at stable weight and there should be an indication of Stable weight.</p> <p>The display should shut off when idle and lights up when weighing operations resume.</p> <p>The system should have RS 232 bi-directional communication port for Data Interface.</p> <p>The system should be CE/ ISO Certified.</p> <p>The system should be supplied with all the accessories required to function.</p> <p>The system should be installed by the trained Engineer and users should be trained on application, operation and maintenance.</p> <p>The product should be as per CE/IEC guideline and certificate from authorized body should be submitted. No self-declaration will be accepted.</p> <p>Compliance to each of the above points should be separately indicated and evidence presence for each of them (Product brochures should be highlighted whenever required).</p> <p>The service team should be in Bangalore. Office address and contact details of Engineer should be submitted with this offer.</p> <p>The past performance/ Service support in NCBS. inStem. CCamp should be satisfactory and the technical evaluation will be done accordingly.</p> <p>The system should have warranty period of 3 years.</p>	02.00 Nos.

2.	<p><b>Precision Weighing Balance</b></p> <p><b>Specifications:</b></p> <p>The system should have Maximum Weighing capacity of 450 g</p> <p>The system should have LCD Display with the attractive backlight.</p> <p>The system should show minimum display of 0.0001 g</p> <p>The system should have repeatability of 0.0001 g</p> <p>The system should have linearity of 0.002 g</p> <p>The system should have “Tare” function available.</p> <p>The system should have “Hold” function available.</p> <p>The system should have MEMORY function to keep the last weight in memory.</p> <p>The system should have inbuilt rechargeable battery backup for minimum of 8 hrs.</p> <p>The system have Internal Calibration option.</p> <p>The system should have response time of min 5 seconds.</p> <p>The system should be compatible in operating at 4°C to 40°C</p> <p>The system should be able to operate at 230V, 50 Hz.</p> <p>The system have option for power saving as a inbuilt feature.</p> <p>The system should have Low Battery indication should be available while operating on battery.</p> <p>The reading should get locked automatically at stable weight and there should be an indication of Stable weight.</p> <p>The display should shut off when idle and lights up when weighing operations resume.</p> <p>The system should have RS 232 bi-directional communication port for Data Interface.</p> <p>The system should be CE/ ISO Certified.</p> <p>The system should be supplied with all the accessories required to function.</p> <p>The system should be installed by the trained Engineer and users should be trained on application, operation and maintenance.</p> <p>The product should be as per CE/IEC guideline and certificate from authorized body should be submitted. No self-declaration will be accepted.</p> <p>Compliance to each of the above points should be separately indicated and evidence presence for each of them (Product brochures should be highlighted whenever required).</p> <p>The service team should be in Bangalore. Office address and contact details of Engineer should be submitted with this offer.</p> <p>The past performance/ Service support in NCBS, inStem, CCamp should be satisfactory and the technical evaluation will be done accordingly.</p> <p>The system should have warranty period of 3 years.</p>	02.00 Nos.
	<p><b>NOTE :1. Please provide authorization letter from OEM along with quotation.</b></p> <p><b>2. The Price quoted against this RFQ should be extended to Bangalore Life Science Cluster (BLiSc) of NCBS, InStem and C-CAMP for placing repeat order as per norms, by any one or all institutes of Bangalore Life Science Cluster (BLiSc)</b></p>	

**1. The quotation shall be submitted in a sealed envelope duly superscribed with the enquiry number, and the due date for ..... Ref No: 192017. The bids should be addressed and to be mailed/couriered (sent by post/courier) to 'THE HEAD-PURCHASE'.**

The bids are liable to be rejected if the sealed envelope is not addressed to “THE HEAD-PURCHASE” with Tender Ref No. and Item Description and due date. The bids delivered in person shall be dropped in Purchase Section. If the bids are sent through courier or mail, it should reach by submission date and time and NCBS will not be responsible for the delay.

**2. DUE DATE FOR SUBMISSION OF QUOTATION AGAINST THIS ENQUIRY IS 10/02/2020 BY 5.30PM.**

3. QUOTATIONS RECEIVED AFTER THE DUE DATE SHALL BE REJECTED.
4. The validity of your quotation should be for 60 days from the due date.
5. All duties, taxes, surcharge and cess as currently applicable must be stated in your quotation, separately. Otherwise your quote is liable to be rejected.
6. Your quotation should indicate delivery period & warranty period.
7. Delivery to be made to our stores. Please indicate charges, if any extra. Transit Insurance should be done upto NCBS Stores.
8. If you are unable to supply the quality, specifications or brand as mentioned in our enquiry, please state so and then offer alternative to quality/specifications.
9. Payment: within one month after delivery & acceptance/satisfactory installation.
10. Please ensure that the enquiry number and the due date is superscribed on the envelope failing which your quotation is liable to be rejected.
11. Since we are a public funded research institution, we are exempted from paying Customs duty (Except ad valorem duty of 5% + 2% Cess and 1% Cus Sec & High Edu. CESS vide Notification No. 51/96 with latest amendments) and excise duty vide Notification No. 10/97 CENTRAL EXCISE dated 01-03-1997 for all scientific equipments, technical instruments, equipments (including computers), their accessories, spares, consumables and software. Hence, please offer your prices.
- 12. If the item is covered under DGS&D rate contract, please quote the rate as per the DGS&D rate contract with xerox copy of the DGS&D order.**
13. Any dispute or differences that may arise between the parties shall be referred to the sole arbitration of the Centre Director or his nominees. The decision of the arbitrator shall be final and binding on the parties. The venue for arbitration shall be Bangalore. The provisions of the Arbitration and Conciliation Act, 1996 as amended from time to time shall apply. The courts in Bangalore shall have exclusive jurisdiction to deal with any or all disputes between the parties.
- 14. The quotations may be submitted in USD, Euro, GBP, JPY and any tradable currency.**
- 15. TIFR-NCBS is a public funded research institute and is entitled to concessional rate of GST @ 5% for certain items supplied for research purpose vide notification no. 47/2017 and 45/2017 dated 14th Nov, 2017. The offer should be submitted after fully considering the above notification.**
16. Liquidity Damages: If the equipment/ items as per specifications in our P.O. is not supplied (shipped) within the specified delivery schedule, then liquidated damages (not in terms of penalty) will be imposed automatically and shall be deducted from the bill at the rate of 0.5% per week subject to a maximum of 10% of the order value.
17. Income Tax at the applicable rates as per the Indian Income Tax Act 1961 will be deducted at source for the services availed / ordered. In case of service provider, the rate of tax deduction shall be at 2% as per Section 194C, and in case of fee for professional / technical services.

Yours faithfully  
For and on behalf of  
National Centre for Biological Sciences



Head Purchase